



CENTRAL UNIVERSITY OF SOUTH BIHAR

S4-7, Gaya- Patanchang Road, Village: Kachua, Post: Patanchang,
P. S. Tazari, District: Gaya (Bihar) PIN- 824 238

CUSBPSC/ABSC/UP/TENDER/2019-20

Date: 06/02/2019

Addendum Notice

Tender Ref. No. : CUSBPSC/ABSC/UP/TENDER/2019-20, Dated: 23/02/2019.
Subject: Supply & Installation of "Laboratory Equipments".

An addendum w.r.t. above mentioned tender notice is being published, for specification mentioned below for Biosafety Cabinet is annexed in Annexure 'Spec-BTW' on tender document page no. 29 of 31.

Sl. No.	Equipment	Specification
1	Biosafety Cabinet	Qualification Working Level II Should meet NSF-49 (2002), EN 12469 (2000) standards Should have double-layer side panels with intrinsic negative pressure Digital display to control of all parameters and safety functions with alarms. Auto blower speed adjustment for filter loading compensation. HEPA down flow and exhaust filter (0.3 - 0.2 um) at 99.999% efficiency Downflow/inflow velocity (m/s) 0.35/0.25/0.22 Flow efficiency with 70% down flow and 30% out flow (checked) UV light & Fluorescent lamp (2 Nos., 110W each) Fluorescent lighting, 1hr light with 3 years warranty Internal dimensions (mm): 900 - 900W x 500 - 600D x 600 - 700H Noise (dB) < 60 Rated voltage 220V - 240 V AC Warranty of 3 years for all parts & labor.

(Anil Kumar)
Assistant Registrar