



# CENTRAL UNIVERSITY OF SOUTH BIHAR

SH-7, Gaya- Panchanpur Road, Village- Karhara, Post- Fatehpur,  
P.S- Tekari, District- Gaya (Bihar) PIN- 824 236

CUSB/PSD/LABEQUIP/TENDER/21/2018-19

Date:05/03/2019

## Addendum Notice

Tender Ref. No. : CUSB/PSD/LABEQUIP/TENDER/21/2018-19, Dated: 22/02/2019.

Subject: Supply & Installation of "Laboratory Equipments".

An addendum w.r.t. above mentioned tender notice is being published, for specification mentioned below for **Biosafety Cabinet** is annexed in **Annexure 'Spec-BTN'** on tender document page no. 29 of 81.

Sl. No.	Equipment	Specification
1	Biosafety Cabinet	Qualifies BioSafety Level II Should meet NSF49:2002, EN12469:2000 standards Should have double-layer side panels with intrinsic negative pressure Digital display to control of all parameters and safety functions with alarms. Auto blower speed adjustment for filter loading compensation. HEPA down flow and exhaust filter (0.1 ~ 0.2 $\mu$ m) at 99.9999% efficiency Downflow/Inflow velocity (m/s): 0.35/0.53 $\pm$ 0.025 Flow efficiency with 70% down flow and 30% out flow (exhaust). UV light & Fluorescent lamp (2 Nos, 1100 lux) fluorescent lighting, Uv light with 3 years warranty Internal dimensions (mm): 800 - 900W x 500 - 600D x 600 - 700H Noise (db): $\leq$ 60 Rated voltage: 220V ~ 240 V AC Warranty of 3 years for all parts & labour.

(Amit Kumar)  
Assistant Registrar